



**Continuous Air Monitoring System (CAMS) Photoionization Detector Concentration Data (15-min Net Avg)**

3/13/2023 Time	Current Action Level ppbv	CAMS 1 VOC ppmv	CAMS 1 VOC ppbv	3/13/2023 Time	Current Action Level ppbv	CAMS 2 VOC ppmv	CAMS 2 VOC ppbv	3/13/2023 Time	Current Action Level ppbv	CAMS 3 VOC ppmv	CAMS 3 VOC ppbv	3/13/2023 Time	Current Action Level ppbv	CAMS 4 VOC ppmv	CAMS 4 VOC ppbv	Notes
0:00	56	0	0	0:00	63	0	0	0:00	N/A	0	0	0:00	48	0	0	
0:15	56	0	0	0:15	63	0	0	0:15	N/A	0	0	0:15	48	0	0	
0:30	56	0	0	0:30	63	0	0	0:30	N/A	0	0	0:30	48	0	0	
0:45	56	0	0	0:45	63	0	0	0:45	N/A	0	0	0:45	48	0	0	
1:00	56	0	0	1:00	63	0	0	1:00	N/A	0	0	1:00	48	0	0	
1:15	56	0	0	1:15	63	0	0	1:15	N/A	0.0261074	26.107444	1:15	48	0	0	
1:30	56	0	0	1:30	63	0	0	1:30	N/A	0.0309171	30.917076	1:30	48	0	0	
1:45	56	0	0	1:45	63	0	0	1:45	N/A	0.0048098	4.8097698	1:45	48	0	0	
2:00	56	0	0	2:00	63	0	0	2:00	N/A	0.0240412	24.041221	2:00	48	0	0	
2:15	56	0	0	2:15	63	0	0	2:15	N/A	0.0048133	4.813291	2:15	48	0	0	
2:30	56	0	0	2:30	63	0	0	2:30	N/A	0	0	2:30	48	0	0	
2:45	56	0	0	2:45	63	0	0	2:45	N/A	0	0	2:45	48	0	0	
3:00	56	0	0	3:00	63	0	0	3:00	N/A	0.0240446	24.04457	3:00	48	0	0	
3:15	56	0	0	3:15	63	0	0	3:15	N/A	0.0048133	4.8132737	3:15	48	0	0	
3:30	56	0	0	3:30	63	0	0	3:30	N/A	0	0	3:30	48	0	0	
3:45	56	0	0	3:45	63	0	0	3:45	N/A	0	0	3:45	48	0	0	
4:00	56	0	0	4:00	63	0	0	4:00	N/A	0.0261044	26.104352	4:00	48	0	0	
4:15	56	0	0	4:15	63	0	0	4:15	N/A	0.0048112	4.8111953	4:15	48	0	0	
4:30	56	0	0	4:30	63	0	0	4:30	N/A	0	0	4:30	48	0	0	
4:45	56	0	0	4:45	63	0	0	4:45	N/A	0	0	4:45	48	0	0	
5:00	56	0	0	5:00	63	0	0	5:00	N/A	0	0	5:00	48	0	0	
5:15	56	0	0	5:15	63	0	0	5:15	N/A	0.0261067	26.106722	5:15	48	0	0	
5:30	56	0	0	5:30	63	0	0	5:30	N/A	0.0048143	4.8143044	5:30	48	0	0	
5:45	56	0	0	5:45	63	0	0	5:45	N/A	0	0	5:45	48	0	0	
6:00	56	0	0	6:00	63	0	0	6:00	N/A	0	0	6:00	48	0	0	
6:15	56	0	0	6:15	63	0	0	6:15	N/A	0	0	6:15	48	0	0	
6:30	56	0	0	6:30	63	0	0	6:30	N/A	0	0	6:30	48	0	0	
6:45	56	0	0	6:45	63	0	0	6:45	N/A	0	0	6:45	48	0	0	
7:00	56	0	0	7:00	63	0	0	7:00	N/A	0	0	7:00	48	0	0	
7:15	56	0	0	7:15	63	0	0	7:15	N/A	0	0	7:15	48	0	0	
7:30	56	0	0	7:30	63	0	0	7:30	N/A	0	0	7:30	48	0	0	
7:45	56	0	0	7:45	63	0	0	7:45	N/A	0	0	7:45	48	0	0	
8:00	56	0	0	8:00	63	0	0	8:00	N/A	0	0	8:00	48	0	0	
8:15	56	0	0	8:15	63	0	0	8:15	N/A	0	0	8:15	48	0	0	
8:30	56	0	0	8:30	63	0	0	8:30	N/A	0	0	8:30	48	0	0	
8:45	56	0	0	8:45	63	0	0	8:45	N/A	0.0611877	61.187652	8:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
9:00	56	0	0	9:00	63	0	0	9:00	N/A	0.0174857	17.48574	9:00	48	0	0	
9:15	56	0	0	9:15	63	0	0	9:15	N/A	0	0	9:15	48	0	0	
9:30	56	0	0	9:30	63	0	0	9:30	N/A	0	0	9:30	48	0	0	
9:45	56	0	0	9:45	63	0	0	9:45	N/A	0.0611881	61.188089	9:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
10:00	56	0	0	10:00	63	0	0	10:00	N/A	0.078673	78.673013	10:00	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
10:15	56	0	0	10:15	63	0	0	10:15	N/A	0.078673	78.673016	10:15	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
10:30	56	0	0	10:30	63	7.332E-06	0.0073323	10:30	N/A	0.078673	78.673014	10:30	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced

10:45	56	0.000167	0.1669979	10:45	63	0.0077446	7.744564	10:45	N/A	0.078673	78.673036	10:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
11:00	56	0.001077	1.0769739	11:00	63	0.0001838	0.1837829	11:00	N/A	0.0174838	17.483817	11:00	48	0	0	
11:15	56	0.002753	2.7530285	11:15	63	0	0	11:15	N/A	0.1182669	118.26689	11:15	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
11:30	56	0.0035302	3.5302255	11:30	63	1.705E-07	0.0001705	11:30	N/A	0.0439294	43.929367	11:30	48	0	0	
11:45	56	0.0041653	4.1653133	11:45	63	0.0001803	0.1802878	11:45	N/A	0	0	11:45	48	0	0	
12:00	56	0.0041886	4.1885913	12:00	63	4.276E-06	0.0042763	12:00	N/A	0	0	12:00	48	0	0	
12:15	56	0.0041674	4.1673724	12:15	63	0	0	12:15	N/A	0	0	12:15	48	0	0	
12:30	56	0.0044153	4.4152552	12:30	63	0	0	12:30	N/A	0	0	12:30	48	0	0	
12:45	56	0.0044922	4.4921576	12:45	63	0	0	12:45	N/A	0	0	12:45	48	0	0	
13:00	56	0.0042543	4.254346	13:00	63	0	0	13:00	N/A	0	0	13:00	48	0	0	
13:15	56	0.0038508	3.8508319	13:15	63	0	0	13:15	N/A	0	0	13:15	48	0	0	
13:30	56	0.0039252	3.9251868	13:30	63	0.0001136	0.1136371	13:30	N/A	0	0	13:30	48	0	0	
13:45	56	0.0041059	4.1058884	13:45	63	1.196E-05	0.0119643	13:45	N/A	0.1358886	135.8886	13:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
14:00	56	0.0045483	4.5483274	14:00	63	0	0	14:00	N/A	0.1857847	185.78472	14:00	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
14:15	56	0.0041859	4.1859009	14:15	63	0	0	14:15	N/A	0.1966944	196.69444	14:15	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
14:30	56	0.003936	3.9359827	14:30	63	0	0	14:30	N/A	0.1991645	199.16453	14:30	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
14:45	56	0.0040233	4.0232881	14:45	63	0	0	14:45	N/A	0.197598	197.59799	14:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
15:00	56	0.0040166	4.0166242	15:00	63	0	0	15:00	N/A	0.0633866	63.386641	15:00	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
15:15	56	0.0042494	4.2494298	15:15	63	0	0	15:15	N/A	0.0025601	2.560057	15:15	48	0	0	
15:30	56	0.004247	4.2469948	15:30	63	0	0	15:30	N/A	0	0	15:30	48	0	0	
15:45	56	0.0039397	3.9396886	15:45	63	0	0	15:45	N/A	0	0	15:45	48	0	0	
16:00	56	0.0041425	4.1424522	16:00	63	0	0	16:00	N/A	0	0	16:00	48	0	0	
16:15	56	0.0056066	5.6066366	16:15	63	0	0	16:15	N/A	0	0	16:15	48	0	0	
16:30	56	0.0069706	6.9706322	16:30	63	0	0	16:30	N/A	0	0	16:30	48	0	0	
16:45	56	0.0072746	7.2745678	16:45	63	0	0	16:45	N/A	0	0	16:45	48	0	0	
17:00	56	0.0060853	6.0853123	17:00	63	0	0	17:00	N/A	0	0	17:00	48	0	0	
17:15	56	0.0049574	4.9573566	17:15	63	0	0	17:15	N/A	0	0	17:15	48	0	0	
17:30	56	0.0041889	4.1889394	17:30	63	0	0	17:30	N/A	0	0	17:30	48	0	0	
17:45	56	0.002775	2.7749641	17:45	63	0	0	17:45	N/A	0	0	17:45	48	0	0	
18:00	56	0.000932	0.9319519	18:00	63	0	0	18:00	N/A	0	0	18:00	48	0	0	
18:15	56	0	0	18:15	63	0	0	18:15	N/A	0	0	18:15	48	0	0	
18:30	56	0	0	18:30	63	0	0	18:30	N/A	0	0	18:30	48	0	0	
18:45	56	6.919E-05	0.0691867	18:45	63	0	0	18:45	N/A	0	0	18:45	48	0	0	
19:00	56	0.0002236	0.2236474	19:00	63	0	0	19:00	N/A	0	0	19:00	48	0	0	
19:15	56	0.0001409	0.1408877	19:15	63	0.0001397	0.1396676	19:15	N/A	0	0	19:15	48	0	0	
19:30	56	0	0	19:30	63	3.991E-05	0.0399113	19:30	N/A	0	0	19:30	48	0	0	
19:45	56	0	0	19:45	63	0	0	19:45	N/A	0.0799555	79.955544	19:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
20:00	56	0	0	20:00	63	0.0003196	0.3196449	20:00	N/A	0.2326808	232.68078	20:00	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
20:15	56	0	0	20:15	63	0.0004039	0.4038748	20:15	N/A	0.3036401	303.64014	20:15	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
20:30	56	0	0	20:30	63	0.0001221	0.1220547	20:30	N/A	0.3300858	330.0858	20:30	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
20:45	56	0	0	20:45	63	3.104E-06	0.0031044	20:45	N/A	0.3527383	352.73831	20:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
21:00	56	0	0	21:00	63	0.0001138	0.1137579	21:00	N/A	0.3738705	373.87047	21:00	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
21:15	56	0	0	21:15	63	4.622E-05	0.0462157	21:15	N/A	0.3886942	388.69425	21:15	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
21:30	56	0	0	21:30	63	0	0	21:30	N/A	0.4049501	404.95005	21:30	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
21:45	56	0	0	21:45	63	0	0	21:45	N/A	0.4190104	419.01044	21:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
22:00	56	0	0	22:00	63	2.05E-05	0.020499	22:00	N/A	0.4338371	433.83712	22:00	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
22:15	56	0	0	22:15	63	2.806E-05	0.0280556	22:15	N/A	0.4497302	449.73023	22:15	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
22:30	56	0	0	22:30	63	8.015E-06	0.0080149	22:30	N/A	0.462137	462.13703	22:30	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
22:45	56	0	0	22:45	63	0.000211	0.2109906	22:45	N/A	0.4710617	471.06175	22:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
23:00	56	0	0	23:00	63	0.0001298	0.1298474	23:00	N/A	0.482684	482.68401	23:00	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
23:15	56	5.138E-05	0.0513819	23:15	63	0.0002373	0.237341	23:15	N/A	0.493134	493.13398	23:15	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
23:30	56	0.0004118	0.4118193	23:30	63	0.0004033	0.4032592	23:30	N/A	0.5020478	502.0478	23:30	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
23:45	56	0.0010049	1.0048799	23:45	63	0.0005676	0.5675537	23:45	N/A	0.5091467	509.14671	23:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced

