



**Continuous Air Monitoring System (CAMS) Photoionization Detector Concentration Data (15-min Net Avg)**

3/14/2023 Time	Current Action Level ppbv	CAMS 1 VOC ppmv	CAMS 1 VOC ppbv	3/14/2023 Time	Current Action Level ppbv	CAMS 2 VOC ppmv	CAMS 2 VOC ppbv	3/14/2023 Time	Current Action Level ppbv	CAMS 3 VOC ppmv	CAMS 3 VOC ppbv	3/14/2023 Time	Current Action Level ppbv	CAMS 4 VOC ppmv	CAMS 4 VOC ppbv	Notes
0:00	56	0.0007116	0.7116486	0:00	63	0.0005286	0.528637	0:00	N/A	0.5167098	516.70979	0:00	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
0:15	56	0	0	0:15	63	0.0003424	0.3424035	0:15	N/A	0.5269061	526.90612	0:15	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
0:30	56	0	0	0:30	63	0.0002066	0.2065552	0:30	N/A	0.5386656	538.66564	0:30	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
0:45	56	0	0	0:45	63	0.0002179	0.2178532	0:45	N/A	0.5469125	546.91253	0:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
1:00	56	0	0	1:00	63	0.0001086	0.108615	1:00	N/A	0.5539831	553.98305	1:00	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
1:15	56	0	0	1:15	63	0	0	1:15	N/A	0.5583562	558.35625	1:15	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
1:30	56	0	0	1:30	63	0	0	1:30	N/A	0.1989573	198.95732	1:30	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
1:45	56	0	0	1:45	63	0.0002273	0.2273429	1:45	N/A	0	0	1:45	48	0	0	
2:00	56	0	0	2:00	63	0.0001878	0.1878251	2:00	N/A	0	0	2:00	48	0	0	
2:15	56	0	0	2:15	63	0	0	2:15	N/A	0	0	2:15	48	0	0	
2:30	56	0	0	2:30	63	0	0	2:30	N/A	0	0	2:30	48	0	0	
2:45	56	0	0	2:45	63	7.261E-05	0.0726093	2:45	N/A	0	0	2:45	48	0	0	
3:00	56	0	0	3:00	63	7.041E-05	0.0704109	3:00	N/A	0	0	3:00	48	0	0	
3:15	56	0	0	3:15	63	6.009E-05	0.0600921	3:15	N/A	0	0	3:15	48	0	0	
3:30	56	0	0	3:30	63	4.44E-05	0.0443998	3:30	N/A	0	0	3:30	48	0	0	
3:45	56	0	0	3:45	63	2.162E-05	0.0216248	3:45	N/A	0	0	3:45	48	0	0	
4:00	56	0	0	4:00	63	2.379E-05	0.0237879	4:00	N/A	0	0	4:00	48	0	0	
4:15	56	0	0	4:15	63	9.058E-05	0.0905796	4:15	N/A	0	0	4:15	48	0	0	
4:30	56	0	0	4:30	63	0.0005329	0.532883	4:30	N/A	0	0	4:30	48	0	0	
4:45	56	0	0	4:45	63	0.0005649	0.5648722	4:45	N/A	0	0	4:45	48	0	0	
5:00	56	0	0	5:00	63	8.779E-05	0.0877949	5:00	N/A	0	0	5:00	48	0	0	
5:15	56	0	0	5:15	63	0.0001507	0.1507058	5:15	N/A	0	0	5:15	48	0	0	
5:30	56	0	0	5:30	63	0.0001442	0.1441643	5:30	N/A	0	0	5:30	48	0	0	
5:45	56	0	0	5:45	63	0.0002085	0.2084545	5:45	N/A	0	0	5:45	48	0	0	
6:00	56	0	0	6:00	63	0.0003066	0.3065868	6:00	N/A	0	0	6:00	48	0	0	
6:15	56	0	0	6:15	63	0.0003815	0.3814814	6:15	N/A	0	0	6:15	48	0	0	
6:30	56	0	0	6:30	63	0.0010368	1.0367846	6:30	N/A	0	0	6:30	48	0	0	
6:45	56	0	0	6:45	63	0.0012296	1.2295879	6:45	N/A	0	0	6:45	48	0	0	
7:00	56	0	0	7:00	63	0.0012288	1.2287741	7:00	N/A	0	0	7:00	48	0	0	
7:15	56	0	0	7:15	63	0.0008345	0.834523	7:15	N/A	0	0	7:15	48	0	0	
7:30	56	0	0	7:30	63	0.0007416	0.74159	7:30	N/A	0	0	7:30	48	0	0	
7:45	56	0	0	7:45	63	0.0018244	1.8243843	7:45	N/A	0.0011253	1.1252798	7:45	48	0	0	
8:00	56	0	0	8:00	63	0.0023772	2.3772386	8:00	N/A	0.0014308	1.4308123	8:00	48	0	0	
8:15	56	0	0	8:15	63	0.0024348	2.4348119	8:15	N/A	0.0020319	2.0318773	8:15	48	0	0	
8:30	56	0	0	8:30	63	0.0032461	3.2461102	8:30	N/A	0.0017713	1.771342	8:30	48	0	0	
8:45	56	0	0	8:45	63	0.0043928	4.3928338	8:45	N/A	0.0227233	22.723281	8:45	48	0	0	
9:00	56	0	0	9:00	63	0.0050204	5.0203829	9:00	N/A	0.0162799	16.279949	9:00	48	0	0	
9:15	56	0.0005833	0.5833423	9:15	63	0.0044946	4.4945713	9:15	N/A	0	0	9:15	48	0	0	
9:30	56	0.0005883	0.5882698	9:30	63	0.0032465	3.2465477	9:30	N/A	0	0	9:30	48	0	0	
9:45	56	0.0001458	0.1457787	9:45	63	0.0035659	3.5658893	9:45	N/A	0	0	9:45	48	0	0	
10:00	56	0.0021602	2.1601983	10:00	63	0.0042566	4.2566125	10:00	N/A	0	0	10:00	48	0	0	
10:15	56	0.0012964	1.296432	10:15	63	0.004209	4.2090159	10:15	N/A	0	0	10:15	48	0	0	
10:30	56	0.0033572	3.3571555	10:30	63	0.0044015	4.401494	10:30	N/A	0	0	10:30	48	0	0	

10:45	56	0.0017756	1.7756346	10:45	63	0.0051255	5.1254705	10:45	N/A	0	0	10:45	48	0	0	
11:00	56	0.0002709	0.2708705	11:00	63	0.0054625	5.4625422	11:00	N/A	0	0	11:00	48	0	0	
11:15	56	2.195E-05	0.0219487	11:15	63	0.0055175	5.5175183	11:15	N/A	0	0	11:15	48	0	0	
11:30	56	0	0	11:30	63	0.0052648	5.2648275	11:30	N/A	0	0	11:30	48	0	0	
11:45	56	0	0	11:45	63	0.0053755	5.3755229	11:45	N/A	0	0	11:45	48	0	0	
12:00	56	0	0	12:00	63	0.0050205	5.0205141	12:00	N/A	0	0	12:00	48	0	0	
12:15	56	0	0	12:15	63	0.0050371	5.0371173	12:15	N/A	0.0501481	50.148144	12:15	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
12:30	56	0	0	12:30	63	0.0051115	5.1115471	12:30	N/A	0.1043001	104.30009	12:30	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
12:45	56	0	0	12:45	63	0.0047026	4.7025676	12:45	N/A	0.0538842	53.884199	12:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
13:00	56	0	0	13:00	63	0.0040019	4.0018553	13:00	N/A	0	0	13:00	48	0	0	
13:15	56	0	0	13:15	63	0.0043974	4.3973794	13:15	N/A	0.0682092	68.209163	13:15	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
13:30	56	0	0	13:30	63	0.0049361	4.9360934	13:30	N/A	0.1606303	160.63031	13:30	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
13:45	56	0	0	13:45	63	0.0052144	5.2144418	13:45	N/A	0.2047114	204.71138	13:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
14:00	56	0	0	14:00	63	0.0053057	5.3057008	14:00	N/A	0.2231578	223.15782	14:00	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
14:15	56	0	0	14:15	63	0.0047964	4.796371	14:15	N/A	0.242424	242.42404	14:15	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
14:30	56	0	0	14:30	63	0.0048924	4.8923526	14:30	N/A	0.2714502	271.45017	14:30	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
14:45	56	0	0	14:45	63	0.0052465	5.246482	14:45	N/A	0.2921398	292.13984	14:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
15:00	56	0	0	15:00	63	0.0051158	5.115805	15:00	N/A	0.311428	311.42801	15:00	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
15:15	56	0	0	15:15	63	0.0044321	4.4320894	15:15	N/A	0.3293926	329.39255	15:15	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
15:30	56	0	0	15:30	63	0.0040025	4.0024668	15:30	N/A	0.3449586	344.95861	15:30	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
15:45	56	0	0	15:45	63	0.0048378	4.8378347	15:45	N/A	0.3596691	359.66906	15:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
16:00	56	0	0	16:00	63	0.00409	4.090028	16:00	N/A	0.3727784	372.77844	16:00	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
16:15	56	0	0	16:15	63	0.0042539	4.2539185	16:15	N/A	0.3845519	384.55189	16:15	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
16:30	56	0	0	16:30	63	0.004541	4.541003	16:30	N/A	0.3978605	397.86053	16:30	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
16:45	56	0	0	16:45	63	0.0033714	3.3713947	16:45	N/A	0.4116604	411.66042	16:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
17:00	56	0	0	17:00	63	0.0030299	3.0299109	17:00	N/A	0.4214369	421.43691	17:00	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
17:15	56	0	0	17:15	63	0.0037994	3.7993897	17:15	N/A	0.4313696	431.36962	17:15	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
17:30	56	0	0	17:30	63	0.0033077	3.307692	17:30	N/A	0.4440409	444.04094	17:30	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
17:45	56	0	0	17:45	63	0.0032792	3.2792372	17:45	N/A	0.4558244	455.82436	17:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
18:00	56	0	0	18:00	63	0.0031141	3.1140624	18:00	N/A	0.46379	463.79001	18:00	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
18:15	56	0	0	18:15	63	0.0026388	2.6387786	18:15	N/A	0.4711718	471.17179	18:15	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
18:30	56	0	0	18:30	63	0.0029177	2.9177209	18:30	N/A	0.4783721	478.37215	18:30	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
18:45	56	0	0	18:45	63	0.002532	2.532046	18:45	N/A	0.4875514	487.55139	18:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
19:00	56	0	0	19:00	63	0.0031523	3.1523448	19:00	N/A	0.4969487	496.9487	19:00	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
19:15	56	0	0	19:15	63	0.0038793	3.8793008	19:15	N/A	0.5047273	504.72733	19:15	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
19:30	56	0	0	19:30	63	0.0028355	2.83551	19:30	N/A	0.5122696	512.2696	19:30	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
19:45	56	0	0	19:45	63	0.0031789	3.1789484	19:45	N/A	0.5197539	519.75389	19:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
20:00	56	0	0	20:00	63	0.0036162	3.6162166	20:00	N/A	0.5288152	528.81516	20:00	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
20:15	56	0	0	20:15	63	0.0031347	3.1347139	20:15	N/A	0.537854	537.85401	20:15	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
20:30	56	0	0	20:30	63	0.00435	4.3500329	20:30	N/A	0.5435784	543.57839	20:30	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
20:45	56	0	0	20:45	63	0.004025	4.0249741	20:45	N/A	0.5484561	548.45607	20:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
21:00	56	0	0	21:00	63	0.0046749	4.6749353	21:00	N/A	0.5541247	554.12475	21:00	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
21:15	56	0	0	21:15	63	0.0051381	5.1381093	21:15	N/A	0.5597934	559.7934	21:15	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
21:30	56	0	0	21:30	63	0.005699	5.6989768	21:30	N/A	0.5671848	567.18476	21:30	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
21:45	56	0	0	21:45	63	0.0057806	5.7805516	21:45	N/A	0.5766115	576.61147	21:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
22:00	56	0	0	22:00	63	0.0055275	5.5275343	22:00	N/A	0.5849522	584.95215	22:00	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
22:15	56	0	0	22:15	63	0.0032313	3.2312749	22:15	N/A	0.5913597	591.35971	22:15	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
22:30	56	0	0	22:30	63	0.0020576	2.0576344	22:30	N/A	0.5961512	596.15117	22:30	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
22:45	56	0	0	22:45	63	0.0054316	5.4316366	22:45	N/A	0.6003511	600.35115	22:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
23:00	56	0	0	23:00	63	0.0053496	5.3496146	23:00	N/A	0.6048214	604.82139	23:00	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
23:15	56	0	0	23:15	63	0.0047897	4.7896695	23:15	N/A	0.6089629	608.96292	23:15	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
23:30	56	0	0	23:30	63	0.0058674	5.8673563	23:30	N/A	0.6137062	613.70615	23:30	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced
23:45	56	0	0	23:45	63	0.0026544	2.6544342	23:45	N/A	0.6195235	619.5235	23:45	48	0	0	* CAMS 3 Faulty Analyzer Being Replaced